

INTENDED USE

For the use in microscopic examination of prepared slides from clinical specimens.

SUMMARY AND EXPLANATION

In light microscopy, oil immersion is a technique used to increase the resolution of a microscope. This is achieved by both the objective lens and the specimen being covered in a transparent oil of high refractive index, thereby increasing the numerical aperture of the objective lens.

The study of cells and cellular structure using standard light microscopy is limited in terms of magnification. In microbiological microscopy this is improved with the use of oil immersion lenses that offer magnification up to x1000 for example for bacterial cells, protozoa and fungal elements.

PRINCIPLE

Image quality is of paramount importance and is directly related to the optical properties of the immersion oil used. The use of immersion oil with the same refractive index as the glass used in the manufacture of glass microscopy slides, increases the resolution between the oil immersion lens and the specimen slide by replacing the air between the lens and the slide easing the spread of light at the same speed as in glass hence avoiding image distortion.

Immersion oil has a 1.513 refractive index, equaling that of glass which guarantees clear images. There is no residual fluorescence, making it suitable for both light and fluorescent microscopy. This also limits common phenomena such as achromatism, spherical aberration and field flatness. It is free from Dibutyl Phthalate DBP and has reduced hazards (non teratogenic).

MATERIALS PROVIDED

- PL.396 Immersion Oil (50ml)

MATERIALS REQUIRED BUT NOT PROVIDED

- Glass Slides
- Stains

STABILITY AND STORAGE

Immersion oil should be kept at room temperature in the original container. Product stored under these conditions will be stable until the expiry date shown on the product label.

PRECAUTIONS

- For *In Vitro* Diagnostic Use only.
- For professional use only.
- Directions should be read and followed carefully.
- Do not use beyond the stated expiration dates.
- Microbial contamination may decrease the accuracy of the staining.
- Safety precautions should be taken in handling, processing and discarding all clinical specimens.
- Samples should be processed in the correct containment level conditions.
- Dispose of all material in accordance with local regulations.

TEST PROCEDURE

1. Focus on the specimen using a dry objective lens (X40), making sure that the specimen is central in the view.
2. Move the dry objective to one side and place a drop of immersion oil on top of the specimen. Move the oil immersion lens into place, ensuring that it makes contact with the oil drop.
3. View the specimen, making any fine adjustments to obtain the best image.
4. After use it is important to clean the oil from the lens using a soft tissue.

QUALITY CONTROL PROCEDURE

Internal quality control of the immersion oil must be performed regularly on known reference material. A selection of quality control slides are available; please see Proslide® range.

INTERPRETATION OF RESULTS










Follow laboratory guidelines.

LIMITATIONS

- Only experienced personnel should carry out the interpretation of stained slides
- Read prepared slides as soon as possible after staining. Failure to do so may affect the results.

REFERENCES

- Anderson, N.L., et al. Cumitech 3B; Quality Systems in the Clinical Microbiology Laboratory, Coordinating ed., A.S. Weissfeld. American Society for Microbiology, Washington, D.C.
- Arnold WM, Weaver RH. Quick microtechniques for the identification of cultures. Journal of Laboratory and Clinical Medicine 1948; 33:1334-7.
- Balzevic, D.J. and G.M. Edrer. (1975). Principles of Biochemical Tests in Diagnostic Microbiology. John Wiley & sons, New York, NY. .
- Chapin, K. C., and T.-L. Lauderdale. 2003. Reagents, stains, and media: bacteriology, p. 354-383.
- Cruickshank, R, J. P. Duguid, B. P. Marmion, R.H.A. Swain. The Practice of Medical Microbiology. 12th Edition. V2
- Gurr, E. A Practical Manual of Medical and Biological Staining Technique. 1953
- Hageage, G.J. B.J. Harrington, Use of calcofluor white in clinical mycology, Journal of Laboratory Medicine, 15, 109-12 (1984).
- Isenberg HD, Ed. Clinical microbiology procedures handbook, Vol I Washington, DC: ASM, 1992.
- Jorgensen., et al. Manual of Clinical Microbiology, American Society for Microbiology, Washington, D.C.
- Lennette. Manual of Clinical Microbiology. Published by American Society for Microbiology, Washington, D.C., 1974
- Lowrance, B.L., P. Reich and W.H. Traub. (1969). Journal of Applied Microbiology 17:923-924.
- Murray, P. R. E. J. Baron, J. H. Jorgensen, M. A. Pfaller, and R. H. Tenover (ed.), Manual of clinical microbiology, 8th edition. ASM Press, Washington, DC. 2003
- Neelson, F. 1883. Ein Casuistischer Beitrag zur Lehre von der Tuberkulose. Centralbl. Med. Wiss. 21:497-501.
- Sutter, V.L. and W.T. Carter. (1972). American Journal of Clinical Pathology 58:335-338
- Wacko, R. and J.C. Sherris. (1963). American Journal of Clinical Pathology. 39:429-432.

	= Use by
	= Lot number
	= Catalogue number
	= Manufacturer
	= Authorized Representative in the European Community
	= Contains sufficient for <n> tests
	= In vitro diagnostic medical device
	= Temperature limitation
	= Consult instructions for use

HAZARD IDENTIFICATION

PL.396	Classification (EC 1272/2008)	NC Not Classified
--------	-------------------------------	-------------------



Advena Ltd. Tower Business Centre, 2nd Floor,
Tower Street, Swatar, BKR 4013, Malta.
T: +44(0) 1926 800 153

